

Š.

#### 輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Certificate No.: C093733

## Certificate of Calibration

#### This is to certify that the equipment

Description: Sound Level Meter

Manufacturer: Rion

Model No.: NL-31

Serial No.: 00320533

has been calibrated for the specific items and ranges. The results are shown in the Calibration Report No. C093783.

The equipment is supplied by

Co. Name: Envirotech Services Co.

Address: Shop 6, G/F., Casio Mansion, 209 Shaukeiwan Road, Hong Kong

Date of Issue: 16 July 2009

Certified by: 1 HC Chan



#### 輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Report No.: C093733

### Calibration Report

ITEM TESTED

DESCRIPTION

Sound Level Meter

MANUFACTURER:

Rion

MODEL NO.

NL-31

SERIAL NO.

00320533

TEST CONDITIONS

AMBIENT TEMPERATURE : (23 ± 2)°C

RELATIVE HUMIDITY: (55 ± 20)%

LINE VOLTAGE

#### TEST SPECIFICATIONS

Calibration check

DATE OF TEST: 15 July 2009

JOB NO.: IC09-1740

#### TEST RESULTS

The results apply to the particular unit-under-test only.

All results are within manufacturer's specification.

The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via:

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Rohde & Schwarz Laboratory, Germany
- Fluke Everett Service Center, USA
- Agilent Technologies, USA

Tested by ;

Date: 16 July 3009

The test equipment used for calibration are traceable to the National Standards as specified in this report. This report shall not be reproduced except in full and with prior written approval from this laboratory.

Calibration and Testing Laboratory of Sun Creation Engineering Limited

e/o 4/F. Tsing Shan Wan Exchange Building, I Hing Octane, Tuen Mun. New Tecritories, Hong Kong

Tel: 2927 2606

Fax: 2744 898€

E-mail: callab@sancreation.com

Website: www.suncreation.com

Page 1 of 4



### 輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Report No. : C093733

### Calibration Report

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 24 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
- 2. Self-calibration was performed before the test.
- 3. The results presented are the mean of 3 measurements at each calibration point,
- 4. Test equipment:

Equipment ID

Description

Certificate No.

CL280 CL281 40 MHz Arbitrary Waveform Generator Multifunction Acoustic Calibrator C090024 DC090052

- Test procedure: MA101N.
- 6. Results:
- 6.1 Sound Pressure Level

6.1.1 Reference Sound Pressure Level

1)	U	JT Setting		Applie	I Value	UUT	IEC 60651 Type 1 Spec. (dB)
Range (dB)	Mode	Frequency Weighting	Time Weighting	Level (dB)	Freq. (kHz)	Reading (dB)	
30 - 120	$L_A$	A	Fast	94.00	1	94.2	± 0.7

6.1.2 Linearity

	U	UT Setting		Applied	Value	UUT		
Range (dB)	Mode	Frequency Weighting	Time Weighting	Level (dB)	Freq. (kHz)	Reading (dB)		
30 - 120	$L_{\Lambda}$	A	Fast	94.00	1	94.2 (Ref.)		
				104.00		104.2		
				114.00		114,2		

IEC 60651 Type 1 Spec. : ± 0.4 dB per 10 dB step and ± 0.7 dB for overall different.

6.2 Time Weighting

6.2.1 Continuous Signal

o* .	UL	JT Setting		Applie	d Value	UUT	IEC 60651
. Range (dB)	Mode	Frequency Weighting	Time Weighting	Level (dB)	Freq. (kHz)	Reading (dB)	Type 1 Spec. (dB)
30 - 120	$L_{A}$	Α	Fast	94.00	1	94.2	Ref.
			Slow			94.1	± 0.1

The test equipment used for calibration are traccable to the National Standards as specified in this report.

The report shall not be reproduced except in tall and with prior written approval from this laboratory.



FROM: ENVIROTECH (HK) LTD

### 輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

PHONE NO.: 85225606553

Report No.: C093733

# Calibration Report

6.2.2 Tone Burst Signal (2 kHz)

Tone Durst		JT Setting	All Control	App	lied Value	UUT	IEC 60651
Range (dB)	Mode	Frequency Weighting	Time Woighting	Level (dB)	Burst Duration	Reading (dB)	Type 1 Spec.
20 - 110	T.,	A	Fast	106.00	Continuous	106.0	Ref.
20-110	L <sub>A</sub>	1			200 ms	105.0	-1.0 ± 1.0
	LAmax Y	-	Slow		Continuous	106.0	Ref.
	Lamer	1			500 ms	102.0	-4.1 ± 1.0

#### 6.3 Frequency Weighting

6.3.1 A-Weighting

-weighting	( C C C C C C C C C C C C C C C C C C C	IT Setting	10 100000	Appli	ed Value	UUT	IEC 60651 Type 1
Range (dB)	Mode	Frequency Weighting	Time Weighting	Level (dB)	Freq.	Reading (dB)	Spec, (dB)
30 - 120	$L_{\Lambda}$	A	Fast	94.00	31.5 Hz	55.0	$-39.4 \pm 1.5$
					63 Hz	68.3	$-26.2 \pm 1.5$
			e.		125 Hz	78.3	-16.1 ± 1.0
28					250 Hz	85.7	-8.6 ± 1.0
					500 Hz	91.0	-3.2 ± 1.0
					1 kHz	94.2	Ref.
					2 kH2	95.2	$+1.2 \pm 1.0$
7					4 kHz	94.4	$+1.0 \pm 1.0$
					8 kHz	90.1	-1.1 (+1.5; -3.0)
9)			1		12.5 kHz	83.9	-4.3 (+3.0 ; -6.0)

6.3.2 C-Weighting

4	UL	T Setting		Appl	ied Value	UUT	IEC 60651 Type 1
Range (dB)	Mode	Frequency Weighting	Time Weighting	Level (dB)	Freq.	Reading (dB)	Spec. (dB)
30 - 120	Lo	С	Fast	94.00	31.5 Hz	91.4	$-3.0 \pm 1.5$
	_				63 Hz	93.6	$-0.8 \pm 1.5$
					125 Hz	94.1	-0.2 ± 1.0
					250 H2	94.3	$0.0 \pm 1.0$
			1		500 Hz	94.3	$0.0 \pm 1.0$
					l kHz	94.2	Ref.
					2 kHz	93.9	-0.2 ± 1.0
					4 kHz	92.7	$-0.8 \pm 1.0$
					8 kHz	88.3	-3.0 (+1.5; -3.0)
40	i				12.5 kHz	82.1	-6.2 (+3.0; -6.0)

The rest equipment used  $\lambda_k$  cally attenues traceable to the National Standards as specified in this report. This report shall not be reproduced except to tun and with order window approval from this laboratory.



### 輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Report No.; C093733

## Calibration Report

Time Averaging

:	U	JT Setting			UUT	IEC 60804				
Range (dB)	Mode	Frequency Weighting	Time Weighting	Freq. (kHz)	Burst Duration (ms)	Burst Duty Factor	Burst Lovel (dB)	Equivalent Level (dB)	Reading (dB)	Type 1 Spec, (dB)
20 - 110	LAGQ	A	10 sec.	4	1	1/10	110.0	100	100.3	± 0.5
W						1/10 <sup>2</sup>		90	90.3	± 0.5
12			60 sec.			1/103		80	80.3	± 1,0
			5 min.			1/104		70	70,3	±1,0

Remarks: - Mfr's Spec.: IEC 60651 & IEC 60804 Type 1

- Uncertainties of Applied Value : 94 dB : 31.5 Hz - 125 Hz :  $\pm$  0.35 dB

250 Hz - 500 Hz :  $\pm 0.30 \text{ dB}$ l kHz  $= \pm 0.20 \text{ dB}$ 2 kHz - 4 kHz  $\pm 0.35 \, dB$ 8 kHz  $\pm 0.45 \, dB$ 12.5 kHz  $\pm 0.70 \text{ dB}$ 

104 dB: 1 kHz = 0.10 dB (Ref. 94 dB)114 dB: 1 kHz  $\pm 0.10 \text{ dB (Ref. 94 dB)}$ Burst equivalent level

 $: \pm 0.2 \text{ dB}$  (Ref. 110 dB) continuous sound level)

- The uncertainties are for a confidence probability of not less than 95 %.

The values given in this Calibration Report only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.

Sun Creation Engineering Limited Calibration and Testing Laboratory

Certificate No.: C095683

## Certificate of Calibration

### This is to certify that the equipment

Description: Sound Level Meter

Manufacturer: Rion

Model No.: NL-31

Serial No.: 00983400

has been calibrated for the specific items and ranges. The results are shown in the Calibration Report No. C095683.

The equipment is supplied by

Co. Name: Envirotech Services Co.

Address: Shop 6, G/F., Casio Mansion, 209 Shaukeiwan Road, Hong Kong

Date of Issue: 23 October 2009

Certified by:

K C Lee



#### 耀創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Certificate No.: C093598

### Certificate of Calibration

This is to certify that the equipment

Description: Sound Level Calibrator

Manufacturer: Rion

Model No.: NC-73

Serial No.: 10786708

has been calibrated for the specific items and ranges. The results are shown in the Calibration Report No. C093598.

The equipment is supplied by

Co. Name: Envirotech Services Co.

Address: Shop 6, G/F., Casio Mansion, 209 Shaukeiwan Road, Hong Kong

Date of Issue: 10 July 2009

Certified by: Chen the HC Chan

Certificate No.: C103765

## Certificate of Calibration

#### This is to certify that the equipment

Description: Sound Level Calibrator

Manufacturer: Rion

Model No.: NC-73

Serial No.: 10997142

has been calibrated for the specific items and ranges. The results are shown in the Calibration Report No. C103765.

#### The equipment is supplied by

Co. Name: Envirotech Services Co.

Address: Shop 6, G/F., Casio Mansion, 209 Shaukeiwan Road, Hong Kong

Date of Issue: 13 July 2010

Certified by:

K C Lee